

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinator, utprovare eller Catell.

Användningsområde

Ortosen stödjer och avlastar brukarens tumme, hand, handled och underarm efter skada, operation, kroniska smärttillstånd, ledförlitning eller överbelastningsbesvär. Ortosen kan även användas i förebyggande syfte vid tunga eller monotona rörelsemoment.

Applicering

1. Ortosen har en förböckad volar skena som kan justeras vid behov.
2. Skenorna runt tummen kan vid behov anpassas efter brukarens tumme och hand.
3. Öppna samtliga kardborrband. Placera handen i ortosen enligt bilden.
4. Stäng samtliga kardborrband så att ortosen sitter behagligt. Justera kardborrar kring tummen.



Klipp av överflödigt kardborrband för ett renare och snyggare utseende. Justera kardborrbanden vid behov. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

Stabil Plus: akryl, bomull, nylon, polyester. Produkten är latexfri.

Tvättråd

Slut kardborrbanden och ta ut skenorna ur skenfickorna. Följ tvättråden nedan. Använd gärna tvättpåse. Låt ortosen droptorka.



Måttabell

Strl	*Mått	Strl	*Mått	Strl	*Mått
XS	12 - 14 cm	M	16 - 18 cm	XL	20 - 22 cm
S	14 - 16 cm	L	18 - 20 cm		

*Mått tas runt handleden och ska ses som en vägledning. Denna produkt är endast avsedd att användas på en patient.

Återvinning

Om produkten har en löstagbar skena, tag ur denna och sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.



Manufacturer Catell AB

Örkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

230416:3

Instruction for use - Prisma Thumb



Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace supports the users thumb, hand, wrist and forearm after injury, surgery, chronic pain, osteoarthritis or overuse problems. The brace can also be used for preventive purposes in heavy or repetitive movements.

Application

1. The brace has a preshaped volar splint. It can be adjusted if necessary.
2. The splints around the thumb can if needed be adapted to the shape of the users thumb and hand.
3. Open all hook and loop straps. Place your hand in the brace as illustrated.
4. Close all hook and loop straps so that the brace fits comfortably. Adjust the hook and loop straps around the thumb.



Cut off excess hook and loop straps for a cleaner and neater appearance. Adjust the hook and loop straps when necessary. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

Material

Stabil Plus: acrylic, cotton, nylon, polyester. Free from latex.

Washing instructions

Close the hook and loop straps and remove the splints from the splint pockets. Follow the washing instructions below. Use a washing pouch. Let the brace drip dry.



Size

Strl	*Dimension	Size	*Dimension	Size	* Dimension
XS	12 - 14 cm	M	16 - 18 cm	XL	20 - 22 cm
S	14 - 16 cm	L	18 - 20 cm		

*Measure wrist circumference. Measurements should be seen as guidelines. Single patient use.

Recycling

If the product has a removable splint, take it out and sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.



Manufacturer Catell AB

Örkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

230416:3